



DECONTAMINATION GUIDANCE

Please read [Ultrasound Transducer Decontamination Best Practice Summary](#) by the British Medical Ultrasound Society and the Society and College of Radiographers: Below

- [Best Practice Summary by the AxREM, BMUS and SCoR \(2020\).](#)

-To prevent cross-infection place our wireless ultrasound probes into a sterile sheath.

-The probes can be cleaned with alcohol wipes.

Antergmix AS1 is a suitable system for disinfection of the wireless probe.

https://pfemedical.co.uk/medical/wp-content/uploads/2018/08/antigermix_s1.pdf

If you require any further advice regarding decontamination or disinfection of our wireless ultrasound probes, please do not hesitate to contact our team at info@ballatermedical.com.